

**Institute of Public Health in Ostrava**

Centre of Hygienic Laboratories

CAI Accredited Testing Laboratory No.1393 according to SN EN ISO/IEC 17025:2005

Partyzánské nám. stí 7, 702 00 Ostrava

**TEST REPORT No. 51518/2017**

**Customer :** CRYSTAL COLLOIDALS  
Thomas Alva Edisonweg 3  
6045 GN Roermond  
NL

**Set No. :** 31795  
**Sample Received :** 15.9.2017 15:00  
**Sample Analyzed :** 15.9.2017 - 18.9.2017  
**Ref. No. :** ZU/ZU/01044/2016  
**File No. :** S-ZU/ZU/01044/2016  
**File code :** 4.0.3

**Sample No. :** 103438  
**Sampling date :** 12.9.2017 **Sampling time:** Not mentioned  
**Sample name:** Crystal colloids Lot.nr: Zn120917/Expiration date:03-2018  
**Sample Type :** colloids  
**Sampled by :** Customer  
**Mode of sampling :** Not mentioned

**Results - chemical analysis**

Parameter	Value	Unit	Type	Method used	Uncertainty
Zn	19,3	mg/l	A	SOP OV 201 <sup>2</sup>	±20%

**Notice to sampling :** The sampling itself is not a subject of accreditation.

**Method specification :**

SOP OV 201 ( SN EN ISO 17294-1, SN EN ISO 17294-2)

**Laboratory workplace :**

<sup>(2)</sup> - Analyses performed at Ostrava (Partyzánské nám. 7, 702 00 Ostrava)

Methods in TYPE column: "A" accredited test

< - the result is below the detection limit, > - the result is higher than the value presented

Results deal with tested samples only.

Without a written consent of the laboratory, this Report can be reproduced only complete.

These expanded uncertainties of measurement are obtained by multiplying of standard uncertainty of measurement by extending coefficient k=2 (for confidence level 95%). Uncertainty of sampling not included.

**Head of Hygienic Laboratories Center :** Došká ová Šárka, RNDr.

**Checked by :** Lach Karel, Ing. CSc.

**Completed by :** Lach Karel, Ing. CSc.

**Number of pages :** 1

**Date :** 18.9.2017

Ing. Vladimíra N mcová  
Head of Department of Inorganic Analyses